



**HEADQUARTERS
PHILIPPINE ARMY
OFFICE OF THE PHILIPPINE ARMY BIDS AND AWARDS COMMITTEE 1
Fort Andres Bonifacio, Metro Manila**

07 October 2021

Supplemental / Bid Bulletin No. 2

**MED PABAC 006-21
Procurement of 3 Lots Medical Equipment for 14 units HMMWV Field Ambulance**

Lot 2 -Procurement of Defibrillator with an ABC of PhP19,502,000.00

This Supplemental Bid Bulletin is issued to inform prospective bidders the amendments/inclusions that shall form part of the bidding documents for the procurement of above mentioned project of the PA;

SOBE	
FROM	TO
13 OCTOBER 2021 @ 1: 30 P.M (WEDNESDAY)	15 OCTOBER 2021 @ 1: 30 P.M (FRIDAY)

Reference		Amendments/Inclusion		
	List of Equipment	Technical Data	List of Equipment	Technical Data
Reference: Section VII. Technical Specification	Defibrillator	AED-Manufacturer's standard Defibrillation- Manufacturer's standard ECG Monitoring- up to 12-Lead ECG Waves Display Waveform: Rectilinear Bi-phasic waveform Integrated CPR feedback with voice and visual prompt Real time CPT feedback dashboard in the monitor, displaying the CPR depth and	Defibrillator	AED-Manufacturer's standard Defibrillation- Manufacturer's standard ECG Monitoring- up to 12-Lead ECG Waves Display Waveform: Bi-phasic waveform that uses low-energy, high current capable of external defibrillation Integrated CPR feedback with voice and visual prompt Real time CPT feedback dashboard in the monitor,

	<p>rate during compression.</p> <p>Monitoring Capabilities: ECG, NIBP, SPO2, ETCO2, Temperature, Non-invasive pacing</p> <p>With heart rate display, arrhythmia & SpO2 alarm</p> <p>With Automatic self-test system</p> <p>Certificate of ISO/CE or its equivalent</p> <p>Power source- autovolt with rechargeable battery pack</p> <p>Accessories:</p> <ul style="list-style-type: none"> - two (2) pcs disposable defibrillator pads and adaptor - two (2) pcs external paddles (Adult & Pedia) <p>-Rechargeable Battery Pack</p>		<p>displaying the CPR depth and rate during compression.</p> <p>Monitoring Capabilities: ECG, NIBP, SPO2, ETCO2, Temperature, Non-invasive pacing</p> <p>With heart rate display, arrhythmia & SpO2 alarm</p> <p>With Automatic self-test system</p> <p>Certificate of ISO/CE or its equivalent</p> <p>Power source- autovolt with rechargeable battery pack</p> <p>Accessories:</p> <ul style="list-style-type: none"> - two (2) pcs disposable defibrillator pads and adaptor - two (2) pcs external paddles (Adult & Pedia) <p>-Rechargeable Battery Pack</p>
--	---	--	--

The Supplemental/Bid Bulletin No. 2 shall be considered an integral part of the bidding documents.

Signed
JESUS JEFFREY F GRAPA
 Colonel (GSC) OS PA
 Vice - Chairperson